



oYo-Link<sup>®</sup> LED PX Device

## Product Specification Sheet

<b>Product Name</b>	oYo-Link <sup>®</sup> LED PX Device
<b>Catalog #</b>	AT8001
<b>User Manual:</b>	See the LED PX Device user manual at: <a href="https://alphathera.com/user-manuals">https://alphathera.com/user-manuals</a>
<b>Outside Dimensions</b>	8 1/2 inch x 9 1/2 inch x 4 inch (L x W x H)
<b>Voltage</b>	110/120V AC 60HZ
<b>Fuse</b>	3A
<b>Bulb Type</b>	LED bulb 365 nm (9W x2 standard)
<b>Bulb Base</b>	Bulb Base G23
<b>Power</b>	18W
<b>Timer</b>	60 mins and 120 mins; Use 120 mins (2 hours) to ensure maximum conjugation efficiency
<b>Fan</b>	Yes
<b>Notes</b>	<p>This LED device is only to be used for in-vitro laboratory research and not to be used for human or clinical samples. Avoid working in moist or humid environments. Always disconnect the main power cord before cleaning the unit or replacing the fuse or LED bulbs. <b>WARNING:</b> The blacklight bulbs used in this device emit long wavelength ultraviolet (UV) light (365 nm), which can pose potential health risks including injuries to the skin and eyes. Please exercise caution when the device is running and avoid eye or skin exposure.</p>

For more information, please visit [www.alphathera.com](http://www.alphathera.com) or contact [support@alphathera.com](mailto:support@alphathera.com).

AlphaThera LLC  
Version 1.0 November 2021